



National
Measurement
Office

Gas Meters – Disputed Meter Accuracy

**Analysis of Findings of Gas Meters Disputed
Between 1 January and 31 December 2012.**

Table of Contents

1. Background.....	1
2. Disputed Gas Meter Findings – Key Points.....	2
3. Prescribed Limits of Accuracy	3
4. Domestic Meters	4
5. Commercial Mechanical (Diaphragm) Meters	5
6. All Industrial (Other than Diaphragm) Meters	6
7. Conclusions.....	7

1. Background

- 1.1 This report details the accuracy findings of gas meters submitted for disputed accuracy testing between 1st January 2012 and 31st December 2012.
- 1.2 The statutory responsibility for the metrological performance of electricity and gas meters was transferred from Ofgem to the National Measurement Office (NMO) on 1st April 2009.
- 1.3 All gas meters in Great Britain (GB) have to be approved¹ and verified² by NMO, or another European Member State, before they can be used for billing purposes. Meters must conform to prescribed accuracy limits for the operating life of the instrument.
- 1.4 There are obligations under the Gas Act 1986 for the owners of meters to 'keep all meters in proper order for correctly registering the quantity of gas supplied'. As such meter owners have a duty to ensure meters are accurate for billing purposes and should make arrangements to monitor their populations of meters and, if required, take action if meters no longer perform to the accuracy limits.
- 1.5 In parallel, NMO provides a service to test and certify the accuracy of gas meters when either a consumer or supplier disputes the meter reading. This work is undertaken, under contract, by a NMO appointed meter examiners employed by SGS (UK).

¹ Laboratory testing of a meter, often a prototype, to ensure it operates accurately in all working conditions

² The checking of meters (either singularly or in batches), often at the premises of the manufacturer or repairer, to ensure they are identical to the model that was type approved and conform to the statutory requirements.

2. Disputed Gas Meter Findings – Key Points

2.1 This information within this report is not intended to represent the overall accuracy of gas meters in Great Britain. The following points should be considered in this context:

- the meters tested were initiated by the consumer or the supplier in circumstances where it was suspected that the meter was measuring erroneously – i.e. the sample is not representative of the whole meter population;
- only a relatively small sample of 2026 meters, manufactured by seven manufacturers, was tested. This needs to be considered against the overall population of gas meters in GB that is in excess of 22 million; and
- this report includes the results of meter tests of types already identified for policy replacement where residual action to remove populations, of the same type, is ongoing.

3. Prescribed Limits of Accuracy

3.1 Limits of meter accuracy for meters approved under UK national legislation are prescribed under The Gas (Meters) Regulations 1983 (SI 684:1983) or The Measuring Instruments (EEC Requirements) (Gas Volume Meters) (Amendment) 1996 (SI 319:1996). Unlike most other countries, there is no additional tolerance for meters operating in-service and are required to meet the same limits of accuracy as when new.


3.2 Gas meters are accurate if they meet the following performance requirements:

• Domestic Mechanical:	±2% at both maximum flow rate; and
	±2% at 20% of maximum flow rate.
• Domestic Electronic:	±2% at both maximum flow rate; and
	±2% at 20% of maximum flow rate; and
	±3% at the minimum flow rate.
• Commercial Mechanical (Diaphragm):	±2% at both maximum flow rate; and
	±2% at 20% of maximum flow rate.
• Industrial (other than Diaphragm):	±1% at 20% of maximum flow rate; and
	±2% at the minimum flow rate.

3.3 For MID meters the limits of meter accuracy are prescribed in The Measuring Instruments (Gas Meters) Regulations 2006 (SI 2647:2006)³. However, there are relatively few MID gas meters installed in GB and, to date, none of these have been submitted for dispute testing.

³ <http://www.legislation.gov.uk/uksi/2006/2647/contents/made>


4. Domestic Meters



4.1 Summary

- Meters designed for a maximum flow rate of 6m³/hr.
- 1865 meters disputed.
- 77% of disputed meters were found to be accurate.
- Other failures – includes meters not registering, not passing gas, excessive leakage, operational problems/events, diagnostic resets, incrementing in a no flow state, and battery failures. These faults prevented meter accuracy from being gained.


5. Commercial Mechanical (Diaphragm) Meters



5.1 Summary:

- Meters designed for a maximum flow rate between 6m³/hr and 160m³/hr.
- 145 meters tested.
- 68% of disputed meters were found to be accurate.
- Other failures – includes meters not registering, not passing gas, excessive leakage, etc. These faults prevented the meter accuracy from being obtained.

6. All Industrial (Other than Diaphragm) Meters



6.1 Summary:

- Meters, of all sizes and manufacturers, comprising of Rotary Positive Displacement (RPD) or Turbine based technology as the principle method for volumetric measurement.
- 16 meters disputed.
- 56% of disputed meters were found to be accurate.
- Other failures – includes meters not registering, not passing gas, excessive leakage, etc. These faults prevented the meter accuracy from being gained.

7. Conclusions

- 7.1 NMO does not use this data for any formal purposes other than to initiate action with meter manufacturers and owners when meter examiners identify systematic faults with certain meter types. This report includes the results of meter tests of types already identified for policy replacement where residual action to remove populations, of the same type, is ongoing. It is important that these factors are considered in context with the information provided.
- 7.2 NMO are working with all industry stakeholders with metering responsibility implement the in-service testing (IST) scheme for domestic type gas and electricity meters. This is a statistically based sampling scheme based on the ongoing monitoring of meter accuracy which will enable energy suppliers and asset owners to demonstrate their meter populations conform to the legal requirements. Further information on the IST scheme is available from the NMO website- <http://www.bis.gov.uk/nmo/gas-and-electricity-meters/In-service-testing-of-gas-and-electricity-meters>
- 7.3 Moving forward, NMO intends to publish the findings of meter dispute testing on an annual basis covering the same parameters as reported in this document.
- 7.4 Further information on gas meter accuracy and billing disputes is available from the NMO website: <http://www.bis.gov.uk/nmo/gas-and-electricity-meters/gas-meters-introduction/Gas-meter-accuracy-and-billing-disputes>
- 7.5 Any enquiries regarding this document should be addressed to:

David Moorhouse
(Assistant Director)
National Measurement Office
Stanton Avenue
Teddington
Middlesex
TW11 0JZ

Email: david.moorhouse@nmo.gov.uk

Tel: 0208 943 7272