



Public Health
England

Protecting and improving the nation's health

1502-02: Germitec – Antigermix AS1

Evaluated by the Rapid Review Panel
10 February 2015

Evaluation 2: Basic research and development has been completed and the product may have potential value; the RRP recommends in use evaluations/trials to demonstrate improved efficiency or efficacy in improving infection prevention and control to reduce healthcare associated infections are considered within an active NHS clinical setting

- This product claims to be an UV-C based High Level Disinfection system for external or endocavitary ultrasound transducers without lumens following appropriate manual cleaning
- The submission presented innovation through rapid disinfection and traceability and contained evidence covering a wide range of microorganisms. Future testing was recommended in the NHS against existing available products

Evaluation Disclaimer: The RRP does not formally review cost-effectiveness or implementation barriers during its evaluation of products; these factors may be considered in a separate process. It is the responsibility of the applicant to carry out any further required development, including trials, cost-effectiveness assessments or product implementation within the NHS. All statements comparing the efficiency or efficacy of products are in the context of products available for use to the NHS at the date of evaluation.