



**1508-01: AntiBac Coating Ltd – Antibac Coating**

Evaluated by the Rapid Review Panel  
11 August 2015

**Evaluation Level 5:** Evidence presented does not demonstrate that the product is more efficient or efficacious at improving infection prevention and control interventions to reduce healthcare associated infections than other available products currently in use.

- The product claims to be a copper/nickel antimicrobial coating which can be applied to most surfaces
- Further microbiological scientific evidence with the product as intended to be used in the NHS is required to substantiate product claims and demonstrate a significant improvement over available products

**Evaluation Disclaimer:** The RRP does not formally review cost-effectiveness or implementation barriers during its evaluation of products; these factors may be considered in a separate process. It is the responsibility of the applicant to carry out any further required development, including trials, cost-effectiveness assessments or product implementation within the NHS. All statements comparing the efficiency or efficacy of products are in the context of products available for use to the NHS at the date of evaluation.