



Protecting and improving the nation's health

**Legionella Isolation Scheme
Preliminary Intended Results Notification**

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **G121** Samples: **G121A and G121B**

Intended results:

	G121A	G121B
*cfu L ⁻¹	<25	2.5x10 ⁵
Species	Not applicable	<i>Legionella bozemanii</i> and <i>Legionella pneumophila</i>
Serogroup	Not applicable	Not applicable and Sg1

* colony forming units per litre

Sample contents:

G121A	<i>Enterococcus faecalis</i> , <i>Pseudomonas aeruginosa</i> and <i>Roseomonas aestuarii</i>
G121B	<i>Legionella bozemanii</i> , <i>Legionella pneumophila</i> and <i>Escherichia coli</i>

Laboratory identification numbers of participants who did not return results this distribution:
395, 994, 1009, 1754, 2077 and 2143

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Legionella Isolation Scheme samples will be dispatched during the week beginning 5 April 2021.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:


Margaret Njenga
Laboratory Supervisor

Date: 29 March 2021



0006