



## SARS-CoV-2 inactivation testing: interim report

<b>Report identifier</b>	HCM/CoV2/015/v2
<b>Report date</b>	15 June 2020
Undertaken by High Containment Microbiology, NIS Laboratories, National Infection Service, Public Health England N.B. This is an interim report and may be updated as further results are obtained	

<b>Product/treatment details</b>	
Product/treatment	cobas® Omni LYS reagent
Manufacturer	Roche
Product code	06997538190
Composition of product, as supplied	30-50% Guanidinium thiocyanate 3-5% Dodecyl alcohol, ethoxylated 1-2.5% Dithiothreitol
Manufacturer's recommended ratio of sample to product	No instructions for use as off-board lysis buffer

<b>Sample details</b>	
Sample type tested	Tissue culture fluid containing 5% (v/v) foetal calf serum
Virus strain tested	SARS-CoV-2 England 2
Ratio of spiked virus stock to sample matrix	Not applicable; tissue culture fluid used undiluted

<b>Experimental conditions</b>	
Ratio of sample to product tested	1 volume sample to 1 volume product
Contact times	10 minutes
Temperature of incubation	Room temperature

Brief description of tests performed	<p>Triplicate samples were treated with test buffer for indicated contact time/s or mock-treated in triplicate with an equivalent volume of PBS. All samples were then subjected to a purification step to remove cytotoxic buffer components. PBS-treated samples were subjected to the same purification procedure in parallel.</p> <p><b>Test 1:</b> Purified samples were immediately titrated on Vero E6 cells to establish virus titre. This test is quantitative and reports the titre of virus in each treatment condition in TCID<sub>50</sub> per ml. Reduction in virus titre following treatment is given as the difference between the mean log<sub>10</sub> TCID<sub>50</sub>/ml for treated conditions and the PBS control.</p> <p><b>Test 2:</b> In parallel, purified samples were seeded onto Vero E6 monolayers to amplify any remaining virus over the course of up to four serial passages. Virus amplification over each passage was detected by visual (microscopic) examination of monolayers for cytopathic effect, and confirmed by SARS-CoV-2-specific real-time PCR. This test is qualitative and reports either the presence or absence of virus amplification. This test may detect levels of virus that are below the detection limit of the titration assay (test 1) due to a greater sample plating volume and the opportunity for any virus present to amplify over serial passages.</p>
--------------------------------------	--

<b>Table of results</b>			
Maximum detectable virus reduction in test (log <sub>10</sub> TCID <sub>50</sub> /ml)			3.9
	<b>Test 1: Virus titration post-treatment</b>		<b>Test 2: Passage of samples in cell culture</b>
	Mean virus titre (log <sub>10</sub> TCID <sub>50</sub> /ml)	Mean titre reduction (log <sub>10</sub> TCID <sub>50</sub> /ml)	Virus detected/ Virus not detected
PBS- treated	6.6	-	Virus detected (all sample replicates)
Test buffer-treated	≤2.7	≥3.9	Virus not detected

## Interpretation

Test 1: Treatment with this buffer resulted in  $\geq 3.9$  log reduction in infectious virus titre, the maximum detectable virus reduction in this test. Treated samples were highly toxic to tissue culture cells (even after purification), reducing the maximum titre reduction for this test.

Test 2: No infectious virus has been detected in treated samples after four serial passages in cell culture.

Demonstrating complete inactivation is dependent on the starting titre of virus used for testing. Sample treatments that inactivate virus effectively in our testing may fail to inactivate samples containing higher levels of virus than those evaluated in this study.

This test has been performed on tissue culture fluid containing 5% (v/v) foetal calf serum. The effectiveness of this treatment against SARS-CoV-2 may vary when used to inactivate clinical samples or other types of sample matrix. Any results of inactivation testing using other sample matrices will be released as they become available.

**Inactivation reagents should not be assumed to be 100% effective against SARS-CoV-2.**

**Suitability of products and treatments for inactivation of other pathogens has not been evaluated in this study.**

**All COVID-19 laboratory testing workflows must be subjected to suitable and sufficient risk assessment, with consideration given to any inactivation step. Risk assessments should be reviewed regularly as new information on the inactivation of SARS-CoV-2 becomes available.**

**The impact of chosen inactivation method on the sensitivity of subsequent SARS-CoV-2 detection should also be assessed locally.**

**Disclaimer**

PHE's evaluations of commercial products and treatments for inactivating SARS-CoV-2 have been carried out primarily for PHE's own internal use and the reports of such evaluations are shared solely for readers information; PHE does not in any way recommend any particular product for virus inactivation; and PHE shall not be responsible for the choice of product or treatment for virus inactivation, and it is the responsibility of the testing laboratory to ensure that any such product or treatment implemented has undergone the necessary verification and validation; and PHE shall not be liable, to the greatest extent possible under any applicable law, for any claim, loss or damage arising out of or connected with use of this and related reports and choice of virus inactivation products or treatments.

PHE is an Executive Agency of the Department of Health and Social Care. Unauthorised use of the PHE name and/or logo is prohibited.

**Summary of revisions**

Version 1: New document

Version 2: Reformatted for publication

Queries regarding this report or HCM inactivation testing should be directed to  
[HCMgroup@phe.gov.uk](mailto:HCMgroup@phe.gov.uk)