

ORGENTEC Diagnostika GmbH • Postfach 100352 • 55134 Mainz

LAUNCH Diagnostics Ltd.  
 Ash House, Ash Road, New Ash Green, Longfield  
 DA3 8JD Kent  
 GREAT BRITAIN

23.05.2013

**Letter of Notification**  
**Anti-dsDNA IgG (ORG 204G), batches 204G30204 and 204G30205**

It has come to our attention that the above batches of Alegria® Anti-dsDNA IgG may show false positive results. We would like to emphasise that this observation is limited to few kits of these batches and the majority of the kits of the above batches fulfill all quality control criteria and are not conspicuous when tested with our quality control samples.

We advise you consider checking to see if you have any results which were unexpectedly positive, that is do not agree with other results or clinical signs. If you think you will need to retest then please contact Launch Diagnostics regarding replacement kits.

Please accept our apologies for the inconvenience that this may have caused you.

With best regards,

**ORGENTEC GmbH**



Jörg Ruppert