

# Medical Device Alert

Ref: MDA/2013/040 Issued: 17 June 2013 at 14:00

Device
<p>Laboratory test kit.</p> <p>Orgentec Alegria dsDNA IgG (ORG204G) for the diagnosis of lupus.</p> <p>Lot numbers: 204G30204 and 204G30205.</p> <p>Manufactured by Orgentec Diagnostika GmbH, distributed in the UK by Launch Diagnostics.</p>

Problem	Action
<p>An increase in sensitivity may lead to false positive results in tests for the diagnosis of lupus. This issue may be related to the strip coating. Orgentec has sent <a href="#">a recall letter</a>.</p>	<p>Do not use and discard any remaining strips from the affected lots.</p> <p>Consider the need to retest positive results obtained from the affected lots.</p> <p>Contact Launch Diagnostics (01474 876 402) to request replacement strips for retesting.</p>
Action by	
<p>Clinical pathologists. Biomedical scientists.</p>	
CAS deadlines	Contact
<p>Action underway: 24 June 2013</p> <p>Action complete: 17 July 2013</p> <p><b>Note: These deadlines are for the removal of unused affected devices and not for the completion of patient re-testing.</b></p>	<p><b>Distributor</b> Ms Louise Knight Launch Diagnostics Ltd Tel: 01474 876 402 Email: <a href="mailto:louiseknight@launchdiagnostics.com">louiseknight@launchdiagnostics.com</a></p>

## Distribution

This MDA has been sent to:

- Care Quality Commission (CQC) (headquarters) for information
- Clinical commissioning groups (CCGs)
- HSC trusts in Northern Ireland (chief executives)
- Local authorities in Scotland (equipment co-ordinators)
- NHS boards and trusts in Wales (chief executives)
- NHS boards in Scotland (equipment co-ordinators)
- NHS trusts in England (chief executives)
- Public Health England (for information)

### Onward distribution

Please bring this notice to the attention of relevant employees in your establishment.  
Below is a suggested list of recipients.

### Trusts

CAS and SABS (NI) liaison officers for onward distribution to all relevant staff including:

- Biomedical scientists
- Clinical governance leads
- Clinical pathologists
- Clinical pathology directors
- Medical directors
- Purchasing managers

### Public Health England

Directors for onward distribution to:

- Laboratory managers

### Independent distribution

#### Establishments registered with the Care Quality Commission (CQC) (England only)

This alert should be read by:

- Hospitals in the independent sector

Please note: CQC and OFSTED do not distribute these alerts. Independent healthcare providers and social care providers can sign up to receive MDAs directly from the Department of Health's Central Alerting System (CAS) by sending an email to: [safetyalerts@dh.gsi.gov.uk](mailto:safetyalerts@dh.gsi.gov.uk) and requesting this facility.

## Contacts

### Manufacturer

Dr Jorg Ruppert  
Orgentec  
Orgentec Diagnostika GmbH  
Carl-Zeiss-Straße 49-51  
55129 Mainz  
Germany

Tel: +49-6131-9258-672

Fax: +49-6131-9258-58

Email: [joerg.ruppert@orgentec.com](mailto:joerg.ruppert@orgentec.com)

**Distributor**

Ms Louise Knight  
Launch Diagnostics Ltd  
Ash House  
Ash Road  
New Ash Green  
Longfield  
DA3 8JD  
UK  
Tel: 01474 876 402  
Email: <mailto:louiseknight@launchdiagnostics.com>

## England

If you are in England, please send enquiries about this notice to the MHRA, quoting reference number **MDA/2013/040** or **2013/005/029/601/004**

**Technical aspects**

Bina Mackenzie or Mojisola Ajeneye  
Medicines & Healthcare Products Regulatory Agency  
Floor 4, 151 Buckingham Palace Road, London SW1W 9SZ  
Tel: 020 3080 7229 or 7271 Fax: 020 8754 3965  
Email: [bina.mackenzie@mhra.gsi.gov.uk](mailto:bina.mackenzie@mhra.gsi.gov.uk)  
[mojisola.ajeneye@mhra.gsi.gov.uk](mailto:mojisola.ajeneye@mhra.gsi.gov.uk)

**Clinical aspects**

**For clinical aspects please contact the clinical team**

Tel: 020 3080 7032 Fax: 020 8754 3965  
Email: [kayleigh.purdon@mhra.gsi.gov.uk](mailto:kayleigh.purdon@mhra.gsi.gov.uk)

### How to report adverse incidents

Please report via our website <http://www.mhra.gov.uk>  
Further information about **CAS** can be found at <https://www.cas.dh.gov.uk/Home.aspx>

## Northern Ireland

Alerts in Northern Ireland will continue to be distributed via the NI SABS system.  
Enquiries and adverse incident reports in Northern Ireland should be addressed to:

Northern Ireland Adverse Incident Centre  
Health Estates Investment Group, Room 17, Annex 6, Castle Buildings, Stormont Estate,  
Dundonald BT4 3SQ

Tel: 02890 523 704 Fax: 02890 523 900 Email: [NIAIC@dhsspsni.gov.uk](mailto:NIAIC@dhsspsni.gov.uk)  
<http://www.dhsspsni.gov.uk/index/hea/niaic.htm>

### How to report adverse incidents in Northern Ireland

Please report directly to NIAIC, further information can be found on our website <http://www.dhsspsni.gov.uk/niaic>  
Further information about **SABS** can be found at <http://sabs.dhsspsni.gov.uk/>

## Scotland

Enquiries and adverse incident reports in Scotland should be addressed to:

Incident Reporting and Investigation Centre

Health Facilities Scotland, NHS National Services Scotland, Gyle Square, 1 South Gyle Crescent,  
Edinburgh EH12 9EB

Tel: 0131 275 7575 Fax: 0131 314 0722 Email: [nss.irc@nhs.net](mailto:nss.irc@nhs.net)

<http://www.hfs.scot.nhs.uk/online-services/incident-reporting-and-investigation-centre-irc/>

## Wales

Enquiries in Wales should be addressed to:

Improving Patient Safety Team

Medical Directorate, Welsh Government, Cathays Park, Cardiff CF10 3NQ

Tel: 029 2082 3922 Email: [Haz-Aic@wales.gsi.gov.uk](mailto:Haz-Aic@wales.gsi.gov.uk)

MHRA is an executive agency of the Department of Health

© Crown Copyright 2013

Addressees may take copies for distribution within their own organisations