

# Medical Device Alert

Ref: MDA/2012/049 Issued: 23 July 2012 at 15:00

## Device

Neonatal and paediatric endotracheal tube clamp/holder.

All product codes.

All lot numbers.

Manufactured by EMS Medical.



Examples of packaging

Problem	Action
<p>Risk of airway loss due to slippage of the clamp/holder</p> <p>or</p> <p>airway obstruction as a result of over-tightening of the clamp/holder.</p> <p>EMS Medical has gone into administration and therefore cannot be contacted.</p> <p>The MHRA has issued this Medical Device Alert to ensure that all users are aware of this issue.</p>	<p>Identify, do not use and dispose of affected devices.</p>
<h3>Action by</h3>	
<p>All staff who may use or handle this device, particularly supplies and paediatric departments.</p>	
<h3>CAS deadlines</h3>	<h3>Contact</h3>
<p>Action underway: 30 July 2012</p> <p>Action complete: 13 August 2012</p> <p><b>Note: These deadlines are for devices to be identified and disposed of.</b></p>	<p><b>No contact available</b></p>

## Device

This device assists in securing an endotracheal tube (ETT) in position for neonatal and paediatric patients.

## Problem

The composition of the plastic used to make the clamp/holder was changed to a DEHP-free material. This affected the physical properties of these components, resulting in an increased risk of slippage or tube obstruction.

EMS Medical went into administration before this issue was identified and it is not known how many affected devices have been supplied.

## Distribution

This MDA has been sent to:

- NHS trusts in England (Chief Executives)
- HSC trusts in Northern Ireland (Chief Executives)
- NHS boards in Scotland (Equipment Co-ordinators)
- Local authorities in Scotland (Equipment Co-ordinators)
- NHS boards and trusts in Wales (Chief Executives)
- Primary care trusts in England (Chief Executives)

### Onward distribution

Please bring this notice to the attention of relevant employees in your establishment. Below is a suggested list of recipients.

### Trusts

CAS and SABS (NI) liaison officers for onward distribution to all relevant staff including:

- A&E departments
- Ambulance staff
- Anaesthetists
- Clinical governance leads
- Day surgery units
- Health and safety managers
- Intensive care medical staff (paediatric)
- Intensive care nursing staff (paediatric)
- Maternity units
- Medical directors
- Midwives
- Neonatal nurse specialists
- Neonatology departments
- Neonatology directors
- Nursing executive directors
- Paediatric surgery, directors of
- Paediatric intensive care units
- Paediatric wards
- Resuscitation officers and trainers
- Risk managers
- Special care baby units
- Supplies managers
- Theatres

### Primary care trusts

CAS liaison officers for onward distribution to all relevant staff including:

- Resuscitation officers

### Independent distribution

#### Establishments registered with the Care Quality Commission (CQC) (England only)

This alert should be read by:

- Hospitals in the independent sector
- Independent treatment centres

Please note: CQC and OFSTED do not distribute these alerts. Independent healthcare providers and social care providers can sign up to receive MDAs directly from the Department of Health's Central Alerting System (CAS) by sending an email to: [safetyalerts@dh.gsi.gov.uk](mailto:safetyalerts@dh.gsi.gov.uk) and requesting this facility.

## England

If you are in England, please send enquiries about this notice to the MHRA, quoting reference number **MDA/2012/049** or **2012/007/005/081/015**.

### Technical aspects

Louise Mulroy and Emma Rooke  
Medicines & Healthcare products Regulatory Agency  
Floor 4  
151 Buckingham Palace Road  
London SW1W 9SZ

Tel: 020 3080 7344 or 6609

Fax: 020 8754 3965

Email: [louise.mulroy@mhra.gsi.gov.uk](mailto:louise.mulroy@mhra.gsi.gov.uk)  
[emma.rooke@mhra.gsi.gov.uk](mailto:emma.rooke@mhra.gsi.gov.uk)

### Clinical aspects

Jonathan Plumb  
Medicines & Healthcare products Regulatory Agency  
Floor 4  
151 Buckingham Palace Road  
London SW1W 9SZ

Tel: 020 3080 6648

Fax: 020 8754 3965

Email: [jonathan.plumb@mhra.gsi.gov.uk](mailto:jonathan.plumb@mhra.gsi.gov.uk)

### How to report adverse incidents

Please report via our website <http://www.mhra.gov.uk>

Further information about **CAS** can be found at <https://www.cas.dh.gov.uk/Home.aspx>

## Northern Ireland

Alerts in Northern Ireland will continue to be distributed via the NI SABS system.

Enquiries and adverse incident reports in Northern Ireland should be addressed to:

Northern Ireland Adverse Incident Centre  
Health Estates Investment Group  
Room 17

Annex 6  
Castle Buildings  
Stormont Estate

Dundonald BT4 3SQ

Tel: 02890 523 704 Fax: 02890 523 900

Email: [NIAIC@dhsspsni.gov.uk](mailto:NIAIC@dhsspsni.gov.uk) <http://www.dhsspsni.gov.uk/index/hea/niaic.htm>

### How to report adverse incidents in Northern Ireland

Please report directly to NIAIC, further information can be found on our website <http://www.dhsspsni.gov.uk/niaic>

Further information about **SABS** can be found at <http://sabs.dhsspsni.gov.uk/>

## Scotland

Enquiries and adverse incident reports in Scotland should be addressed to:

Incident Reporting and Investigation Centre

Health Facilities Scotland

NHS National Services Scotland

Gyle Square

1 South Gyle Crescent

Edinburgh EH12 9EB

Tel: 0131 275 7575 Fax: 0131 314 0722

Email: [nss.irc@nhs.net](mailto:nss.irc@nhs.net)

<http://www.hfs.scot.nhs.uk/online-services/incident-reporting-and-investigation-centre-irc/>

## Wales

Enquiries in Wales should be addressed to:

Improving Patient Safety Team

Medical Directorate

Welsh Government

Cathays Park

Cardiff CF10 3NQ

Tel: 029 2082 3922 Email: [Haz-Aic@wales.gsi.gov.uk](mailto:Haz-Aic@wales.gsi.gov.uk)

MHRA is an executive agency of the Department of Health

© Crown Copyright 2012

Addressees may take copies for distribution within their own organisations