



# M.A. ARAIN & BROTHERS (PVT.) LTD.

Manufacturers & Exporters of fine quality Surgical, Electrosurgical & Diagnostic Instruments

## Urgent Field Safety Notice Laryngoscope Battery Handles 670160-0000\*\*

M. A. Arain Reference Number: MB/509/13/01  
Lot number: 041511 & 021512  
Description of the defective products: (Laryngoscope disposable metallic kit)

### Description of the problem:

During examination of a complaint it was discovered that some coil springs in the battery compartment where improperly engaged during manufacturing. It most likely will be discovered before the application but it may happen during intubation.

### Advise on action to be taken by the user:

Please send back affected products for replacement and check similar other products intend to the same issue before use.

### Transmission of this Field Safety Notice:

This notice needs to be passed on all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred. (If appropriate)

The undersigned confirms that the appropriate Competent Authorities have been advised of the FSCA.

Mr. Waqas Akbar

*Waqas Akbar*

Director Marketing and QA

Date 22-05-2012



C(GMP) FDA, U.S.A.  
CERTIFIED

MEMBERS



MAILING ADDRESS: P.O. BOX 432, SIALKOT-PAKISTAN  
FACTORY: WAZIRABAD ROAD, SIALKOT.  
TEL: +92-52-4264926, 4264571 FAX: +92-52-4264759, 4264026  
E-MAIL: info@maarain.com Web Address: http://www.maarain.com