

# Medical Device Alert


## Action

Ref: MDA/2010/050 Issued: 15 June 2010 at 15:00

## Device

Implantable vagus nerve stimulators (VNS):

- Pulse and Pulse Duo Generators (Models 100, 101, 102 and 102R)
- Demipulse and Demipulse Duo Generators (Models 103 and 104)
- Model 250 programming software.



Manufactured by Cyberonics VNS Therapy.

Problem	Action
<ul style="list-style-type: none"> <li>• Premature battery depletion (models 103 and 104).</li> <li>• Inaccurate battery life projection (models 103, 104 and 250).</li> <li>• Loss of therapy (all models).</li> <li>• Painful, erratic or atypical stimulation (all models).</li> </ul>	<ul style="list-style-type: none"> <li>• Follow the advice in the manufacturer's Field Safety Notices (FSNs) issued in:               <ul style="list-style-type: none"> <li><a href="#">July 2009</a></li> <li><a href="#">November 2009</a></li> <li><a href="#">April 2010</a></li> </ul> </li> <li>• Ensure that you acknowledge each FSN that applies to your organisation by returning the receipt to the manufacturer.</li> </ul>
Action by	
<p>Implanting surgeons and clinicians involved in the management and follow-up of patients with these devices.</p>	
CAS deadlines	Contact
<p>Action underway: 14 July 2010 Action complete: 09 September 2010</p>	<p><b>Manufacturer</b> Cyberonics UK contact: John Forsey Tel: 01246 261 397 Fax: 0870 166 0501 Email for clinical and technical queries: <a href="mailto:europaclintechservices@cyberonics.com">europaclintechservices@cyberonics.com</a></p>

## Problem

The MHRA has issued this Medical Device Alert (MDA) to support Cyberonics' corrective actions, as they have informed us that to date there has been a low level of response to their Field Safety Notices (FSNs) listed below.

In July 2009 Cyberonics issued a [FSN](#) concerning Model 250 Programming software when used with Pulse and Pulse Duo Generators (Models 100, 101, 102 & 102R) and Demipulse and Demipulse Duo Generators (Models 103 and 104).

In November 2009 Cyberonics issued a [FSN](#) concerning the Demipulse and Demipulse Duo Generators (Models 103 and 104).

In April 2010 Cyberonics issued a [FSN](#) concerning Model 250 Programming software when used with Demipulse and Demipulse Duo Generators (Models 103 and 104).

Cyberonics has requested that clinicians contact them as soon as possible if they should have received any of the above FSNs but did not.

It is important to follow the actions advised in FSNs, and for your organisation to acknowledge receipt of FSNs. The receipt provides the manufacturer, and subsequently the MHRA, with the means to monitor the progress of Field Safety Corrective Actions. It also minimises the need for the MHRA to issue MDAs, which otherwise place an additional burden on the health service because of the broadcast nature of the MDA and the extra administrative work required.<sup>1</sup>

<sup>1</sup> [Reporting Adverse Incidents and Disseminating Medical Device Alerts - DB 2010\(01\)](#)

## Distribution

This MDA has been distributed to:

- NHS trusts in England (Chief Executives)
- HSC trusts in Northern Ireland (Chief Executives)
- NHS boards in Scotland (Chief Executives)
- NHS boards and trusts in Wales (Chief Executives)
- Primary care trusts in England (Chief Executives)

### Onward distribution

Please bring this notice to the attention of all who need to know or be aware of it. This may include distribution by:

#### Trusts to:

CAS and SABS (NI) liaison officers for onward distribution to all relevant staff including:

- A&E consultants
- A&E departments
- A&E directors
- A&E nurses
- Adult intensive care units
- All departments
- All staff
- All wards
- Anaesthetists
- Clinical governance leads
- Day surgery units
- Equipment stores
- General surgeons
- General surgery
- General surgical units, directors of
- Health and safety managers
- In-house maintenance staff
- Medical directors
- Medical oncologists
- MRI units, directors of
- Nursing executive directors
- Neurological surgeons
- Neurology and neurosurgery departments
- Occupational therapists

- Oncology nurse specialists
- Outpatient clinics
- Outpatient theatre managers
- Outpatient theatre nurses
- Palliative care teams
- Purchasing managers
- Radiation and medical oncology departments
- Radiation oncology, directors of
- Radiographers
- Radiologists
- Radiology departments
- Radiology directors
- Risk managers
- Supplies managers
- Theatre managers
- Theatre nurses
- Theatres

**Care Quality Commission (CQC) (England only) to:**

The MHRA considers this information to be important to:

- Hospices
- Hospitals in the independent sector
- Independent treatment centres
- Private medical practitioners

**Primary care trusts to:**

CAS liaison officers for onward distribution to all relevant staff including:

- Directors of public health
- Palliative care team
- Practice managers
- Practice nurses

## Contacts

**Manufacturer/Supplier**

John Forsey  
Cyberonics Europe  
Unit 4 Dunston Innovation Centre  
Dunston Road  
Chesterfield  
S41 8NG

Tel: 01246 261 397  
Fax: 0870 166 0501  
Email: [europeclintechservices@cyberonics.com](mailto:europeclintechservices@cyberonics.com)

## England

If you are in England, please send enquiries about this notice to the MHRA, quoting reference number **MDA/2010/050**.

**Technical aspects**

Simon Holmes or Feza Haque  
Medicines & Healthcare products Regulatory Agency  
Market Towers  
1 Nine Elms Lane  
London SW8 5NQ

Tel: 020 7084 3240 / 3066  
Fax: 020 7084 3106  
Email: [simon.holmes@mhra.gsi.gov](mailto:simon.holmes@mhra.gsi.gov)  
[feza.haque@mhra.gsi.gov](mailto:feza.haque@mhra.gsi.gov)

**Clinical aspects**

Dr Susanne Ludgate  
Medicines & Healthcare products Regulatory Agency  
Market Towers  
1 Nine Elms Lane  
London SW8 5NQ

Tel: 020 7084 3123  
Fax: 020 7084 3111  
Email: [susanne.ludgate@mhra.gsi.gov](mailto:susanne.ludgate@mhra.gsi.gov)

### How to report adverse incidents

Please report via our website <http://www.mhra.gov.uk>

Further information about **CAS** can be found at <https://www.cas.dh.gov.uk/Home.aspx>

## Northern Ireland

Alerts in Northern Ireland will continue to be distributed via the NI SABS system.

Enquiries and adverse incident reports in Northern Ireland should be addressed to:

Northern Ireland Adverse Incident Centre

Health Estates Investment Group

Room 17

Annex 6

Castle Buildings

Stormont Estate

Dundonald BT4 3SQ

Tel: 02890 523 704

Fax: 02890 523 900

Email: [NIAIC@dhsspsni.gov.uk](mailto:NIAIC@dhsspsni.gov.uk)

<http://www.dhsspsni.gov.uk/index/hea/niaic.htm>

### How to report adverse incidents in Northern Ireland

Please report directly to NIAIC, further information can be found on our website <http://www.dhsspsni.gov.uk/niaic>

Further information about **SABS** can be found at <http://sabs.dhsspsni.gov.uk/>

## Scotland

Enquiries and adverse incident reports in Scotland should be addressed to:

Incident Reporting and Investigation Centre

Health Facilities Scotland

NHS National Services Scotland

Gyle Square

1 South Gyle Crescent

Edinburgh EH12 9EB

Tel: 0131 275 7575

Fax: 0131 314 0722

Email: [nss.irc@nhs.net](mailto:nss.irc@nhs.net)

<http://www.hfs.scot.nhs.uk/online-services/incident-reporting-and-investigation-centre-irc/>

## Wales

Enquiries in Wales should be addressed to:

Dr Sara Hayes

Senior Medical Officer

Medical Device Alerts

Welsh Assembly Government

Cathays Park

Cardiff CF10 3NQ

Tel: 029 2082 3922

Email: [Haz-Aic@wales.gsi.gov.uk](mailto:Haz-Aic@wales.gsi.gov.uk)

MHRA is an executive agency of the Department of Health

© Crown Copyright 2010

Addressees may take copies for distribution within their own organisations