



Public Health  
England

Protecting and improving the nation's health

**MRRS 4201: Alphasonics UCS Ltd – Ultrasonic Washer Disinfector Medstar Series**

Evaluated by the Rapid Review Panel  
10 February 2015

**Recommendation 5:** insufficient clarity and/or evidence presented to enable a full review of the product.

- This product claims to be an ultrasonic cleaner which removes proteins from soiled surgical instruments
- The submission did not contain sufficient evidence to substantiate disinfection claims or claims of improved performance above appropriate in use comparators

**Evaluation Disclaimer:** The RRP does not formally review cost-effectiveness or implementation barriers during its evaluation of products; these factors may be considered in a separate process. It is the responsibility of the applicant to carry out any further required development, including trials, cost-effectiveness assessments or product implementation within the NHS. All statements comparing the efficiency or efficacy of products are in the context of products available for use to the NHS at the date of evaluation.