



This is to certify that:

Panasonic Eco Solutions Elektrik Sanayi Ticaret Anonim Sirketi Abdurrahmangazi mah Ebubekir cad. No:44 Sancaktepe Istanbul 34887 Turkey

Holds Certificate Number:

0086/BSI/635935

In respect of:

Directive reference: MID 2014/32/EU Annex Item: MI-003 – Active Electrical Energy Meter Product Description: VEM-M580DB0M, VEM-M580DB2M 1 Phase 230V, 0.25-5(80)A, 50Hz, Class B, 3K7, IP54. Software version 1.01.

Specified standard: EN 50470-1:2006, EN 50470-3:2006

on the basis that BSI carried out the relevant Type Examination procedures for the equipment identified above which was found to be in compliance with the Directive 2014/32/EU of the European Parliament and Council relating to the Measuring Instruments Directive (MID), subject to any conditions in the schedule attached hereto. The attached schedule of approval forms part of this certificate.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 0086):

First Issued: 2015-10-07

Gary Fenton, Global Product Certification Director

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 1 of 6

...making excellence a habit."

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/635935

Schedule of Approval

Technical Data

Manufacturer	Panasonic	Meter Type	VEM-M580DB0M & VEM-M580DB2M
Voltage Rating (Un)	230V	Current Rating (Imin-Iref(Imax))	250mA - 5A - 80A
Frequency (Fn)	50Hz	Active Accuracy Class (kWh)	Class B
Type of Circuit	1p2w	Temperature Range	-40 +70
Software Version No. Identification Location CheckSum PCB/Schematic Ref	Ver 1.01 LCD: 09B54 265233R0.1 26/02/15	Bill of Material Nos.	92903720, dated 11/06/15 & 92903901, dated 23/09/16
IP Rating	IP54	Insulation Protective Class	Class II
LED Pulse Constant	3.200 imp/kWh	Impulse Voltage Rating	6kV
AC Voltage Rating	4kV	Seal for EC Initial Verification Type	Plastic Seal on Top Cover and Seal on Terminal cover
Main Cover Seal Type	Push Fit Plastic seal	Integrity of Meter	Inaccessible without breaking the seal
Intended Location of the Meter	Outdoor	Type of Register	LCD

Certification is based on the Turkey accredited laboratory ELDAS, Report number EL2701/R1 dated 15/07/2015, EL16-0019 dated 20/07/16 & NMi Report Number 15200270-01 dated 15/04/15 and Technical Documentation supplied to BSI, project Ref 8335098 & 8624843.

Condition for Conformity Assessment according to Annex D or F

The influence factors for temperature, frequency and voltage, which are necessary to perform the conformity assessment according to annex D of F, are presented on page 3 of this certificate. Based on WELMEC Guide 11.1 Section 2.5.6, the sum of the square values is presented.

During the type approval examination the influence factors for temperature, frequency and voltage are determined per load point. The table below presents the sum of the square values per load, determined via the following formula

$$\begin{split} &\delta e\left(\mathsf{T},\mathsf{U},\mathsf{f}\right) = \sqrt{\delta}e^{2}\left(\mathsf{T},\mathsf{I},\cos\varphi\right) + \delta e^{2}\left(\mathsf{U},\mathsf{I},\cos\varphi\right) + \delta e^{2}\left(\mathsf{f},\mathsf{I},\cos\varphi\right) \\ & \text{Where:} \\ &\delta e\left(\mathsf{T},\mathsf{I},\cos\varphi\right) = \text{the additional error due to the variation of the temperature at a certain load.} \\ &\delta e\left(\mathsf{U},\mathsf{I},\cos\varphi\right) = \text{the additional percentage error due to the variation of the voltage at the same load.} \\ &\delta e\left(\mathsf{f},\mathsf{I},\cos\varphi\right) = \text{the additional percentage error due to the variation of the frequency at the same load.} \end{split}$$

First Issued: 2015-10-07

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 2 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.


To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/635935

			Influence Factor for Temperature, Frequency & Voltage							
	Current	PF Cos	-40°C	-25°C	-10°C	+5°C	+30°C	+40°C	+55°C	+70°C
Single	Imin	1.0	0.40	0.34	0.27	0.26	0.25	0.27	0.34	0.44
Phase	Itr	1.0	0.34	0.23	0.14	0.12	0.10	0.16	0.27	0.39
	10Itr	1.0	0.37	0.31	0.22	0.18	0.18	0.23	0.29	0.36
	Imax	1.0	0.38	0.25	0.18	0.16	0.14	0.19	0.31	0.37
	Itr	0.5ind	0.40	.0.30	0.20	0.15	0.15	0.20	0.28	0.33
	10Itr	0.5ind	0.33	0.27	0.18	0.15	0.13	0.22	0.26	0.35
	Imax	0.5ind	0.35	0.25	0.20	0.13	0.11	0.16	0.25	0.33
	Itr	0.8cap	0.36	0.29	0.21	0.12	0.12	0.18	0.24	0.34
	10Itr	0.8cap	0.34	0.26	0.22	0.16	0.14	0.18	0.26	0.37
	Imax	0.8cap	0.39	0.28	0.23	0.15	0.15	0.18	0.31	0.37
										All and a second second

Schedule of Approval

Seal Locations



First Issued: 2015-10-07

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 3 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request. To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated <u>online</u>.

No. 0086/BSI/635935

Schedule of Approval

Photo



Marking

The following data shall be quoted on the meter name-plate:

- Identification of manufacturer (name or trade mark);
- Type designation;
- "CE" marking and supplementary metrology marking;
- EC-type examination certificate number; _
- Serial number and year of production; _
- _ Class index;
- Specified operating temperature range; _
- Type of distribution network;
- -Reference voltage;
- Reference (rated) current; _
- Maximum current; _
- Minimum current; _
- Reference frequency;
- Constant of meter;
- _ Sign of double-square for meter of protective class II.

First Issued: 2015-10-07

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 4 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request. To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated <u>online</u>.

No. 0086/BSI/635935

Schedule of Approval

Product Variant Identification Details

	METER TYPE CODE TABLE						
PRODUCT CODE	METERING	CURREN T	TYPE OF CONNECTION	METERING CLASS	COMMUNICATIO N OUTPUT	MODEL	
VEM -	Х	XXXXX	Х	Х	Х	Х	
VİKO ENERGY METERING	M = MONO- PHASE	1080	D = DIRECT CONNECTION	A = CLASS A	0= NONE	NULL CHARACTER NORMAL	
	T = THREE- PHASE	10100	C = WITH CURRENT TRANSFORMER	B = CLASS B	1 = RS232	P= PREPAID	
A = MODEL	C = COMBO (INDUSTRIAL)	x5	V = WITH CURRENT/VOLTAGE TRANSFORMER	C = CLASS C	2 = RS485		
		1060 5 (80) 5 (100)			3 = PLC		

Manufacturer Address:

Panasonic Eco Solutions Elektrik Sanayi Ticaret Anonim Sirketi

Abdurrahmangazi Mah Ebubekir Cad. No: 44 34887 Samandıra Istanbul Turkey

Document Revision History

Issue	Date	Comments
1	07/10/2015	Initial Issue.
2	08/11/2016	Minor modification, Software version change - BSI Project Ref 8624843.
3	08/05/2017	Company Name Change

First Issued: 2015-10-07

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 5 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request. To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated <u>online</u>.

No. 0086/BSI/635935

Schedule of Approval

Conditions of Certification

- This 0086/BSI/635935 certificate remains valid unless cancelled or revoked, provided the conditions listed below are i) complied with and the equipment remains satisfactory in service.
- ii) This certificate must be used in conjunction with a certificate covering the product certification as required by Annex D or Annex F.
- Detailed User instructions are to be provided with each product. iii)
- iv) If the specified standards are amended during the validity of this certificate, this product type is to be re-approved.
- Any modifications to the Meter(s) listed on page 1 must be notified to BSI to confirm the Meter(s) continuing v) compliance to the relevant pattern approval standard(s).

First Issued: 2015-10-07

Latest Issue: 2017-05-08

Expiry Date: 2025-10-06

Page: 6 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request. To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated <u>online</u>.