



This is to certify that:

Panasonic Eco Solutions Elektrik Sanayi

Ticaret Anonim Sirketi Abdurrahmangazi mah

Ebubekir cad. No:44 Sancaktepe

Istanbul 34887 Turkey

Holds Certificate Number:

0086/BSI/632148

In respect of:

Directive reference: MID 2014/32/EU Annex Item: MI-003 – Active Electrical Energy Meter

Product Description: VEM-CX5CB2M, 3 Phase 230/400V, 0.05-5(6)A, 50Hz, Class B, 3K7, IP54.

Software version 1.01.

Specified standard: EN 50470-1:2006, EN 50470-3:2006.

on the basis that BSI carried out the relevant Type Examination procedures for the equipment identified above which was found to be in compliance with the Directive 2014/32/EU of the European Parliament and Council relating to the Measuring Instruments Directive (MID), subject to any conditions in the schedule attached hereto. The attached schedule of approval forms part of this certificate.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 0086):

Gary Fenton, Global Product Certification Director

First Issued: 2015-05-12

Latest Issue: 2017-05-08

Expiry Date: 2025-05-11

Page: 1 of 6

...making excellence a habit."

No. 0086/BSI/632148

Schedule of Approval

Technical Data

| Manufacturer | Panasonic | Meter Type | VEM-CX5CB2M |
|---|----------------------------|--|--|
| Voltage Rating (Un) | 3x230/400V | Current Rating (Imin-Iref(Imax)) | 50mA - 5A - 6A |
| Frequency (Fn) | 50Hz | Active Accuracy Class (kWh) | Class B |
| Type of Circuit | 3p4w | Temperature Range | -40 +70 |
| Software Version No. Identification Location CheckSum | Ver 1.01 LCD: 0x1260 | Bill of Material No.'s | 92903709 dated 18/04/2014 & 23/09/16 |
| IP Rating | IP54 | Insulation Protective Class | Class II |
| LED Pulse Constant | 10.000 imp/kWh | Impulse Voltage Rating | 6kV |
| AC Voltage Rating | 4kV | Seal for EC Initial Verification Type | Plastic Seal on Top Cover and Seal on Terminal cover |
| Main Cover Seal Type | Push Fit Plastic seal | Integrity of Meter | Inaccessible without breaking the seal |
| Intended Location of the Meter | Outdoor | Type of Register | LCD |

Certification is based on the Turkey accredited laboratory ELDAS, Report number EL-2102/R1 dated 12/02/2015, EL16-0-22 dated 20/07/16 & NMI Test Report 15200270-01 dated 15/04/2015 and Technical Documentation supplied to BSI, project Ref 8293358 & 8624843

Condition for Conformity Assessment according to Annex D or F

The influence factors for temperature, frequency and voltage, which are necessary to perform the conformity assessment according to annex D of F, are presented on page 3 of this certificate. Based on WELMEC Guide 11.1 Section 2.5.6, the sum of the square values is presented.

During the type approval examination the influence factors for temperature, frequency and voltage are determined per load point. The table below presents the sum of the square values per load, determined via the following formula

$$\delta e (T, U, f) = \sqrt{\delta}e^{2} (T, I, \cos\varphi) + \delta e^{2} (U, I, \cos\varphi) + \delta e^{2} (f, I, \cos\varphi)$$

 $\delta e(T, I, \cos \varphi) = \text{ the additional error due to the variation of the temperature at a certain load}$

 $\delta e (U, I, \cos \varphi) = \text{ the additional percentage error due to the variation of the voltage at the same load}$

 $\delta e(f, I, \cos \phi) = \text{ the additional percentage error due to the variation of the frequency at the same load.}$

First Issued: 2015-05-12 Latest Issue: 2017-05-08 Expiry Date: 2025-05-11

Page: 2 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/632148

Schedule of Approval

| | | | Influence Factor for Temperature, Frequency & Voltage | | | | | | | |
|----------------|------------------------------|--------------------------|---|------------------------------|------------------------------|------------------------------|------------------------------|------------------------------|------------------------------|------------------------------|
| | Current | PF Cos | -40°C | -25°C | -10°C | +5°C | +30°C | +40°C | +55°C | +70°C |
| R S T Phase | Imin Itr 10Itr Imax | 1.0 1.0 1.0 1.0 | 0.41 0.38 0.41 0.40 | 0.31 0.34 0.27 0.32 | 0.29 0.28 0.21 0.25 | 0.25 0.23 0.19 0.54 | 0.25 0.24 0.20 0.23 | 0.29 0.25 0.24 0.31 | 0.34 0.34 0.30 0.36 | 0.45 0.38 0.38 0.41 |
| | Itr | 0.5ind | 0.44 | 0.35 | 0.29 | 0.27 | 0.27 | 0.31 | 0.34 | 0.45 |
| | 10Itr | 0.5ind | 0.58 | 0.53 | 0.52 | 0.50 | 0.49 | 0.50 | 0.54 | 0.64 |
| | Imax | 0.5ind | 0.67 | 0.62 | 0.57 | 0.68 | 0.56 | 0.57 | 0.62 | 0.64 |
| | Itr | 0.8cap | 0.46 | 0.40 | 0.35 | 0.31 | 0.31 | 0.36 | 0.38 | 0.42 |
| | 10Itr | 0.8cap | 0.44 | 0.33 | 0.26 | 0.24 | 0.23 | 0.27 | 0.34 | 0.42 |
| | Imax | 0.8cap | 0.34 | 0.31 | 0.19 | 0.17 | 0.19 | 0.21 | 0.26 | 0.32 |
| R Phase | Itr | 1.0 | 0.38 | 0.30 | 0.24 | 0.20 | 0.21 | 0.23 | 0.33 | 037 |
| | 10Itr | 1.0 | 0.35 | 0.27 | 0.24 | 0.18 | 0.17 | 0.24 | 0.28 | 0.41 |
| | Imax | 1.0 | 0.39 | 031 | 0.23 | 0.16 | 0.16 | 0.21 | 0.26 | 0.34 |
| | Itr | 0.5ind | 0.46 | 0.35 | 0.29 | 0.27 | 0.28 | 0.29 | 0.37 | 0.44 |
| | 10Itr | 0.5ind | 0.56 | 0.52 | 0.48 | 0.46 | 0.46 | 0.50 | 0.52 | 0.54 |
| | Imax | 0.5ind | 0.62 | 0.59 | 0.54 | 0.52 | 0.53 | 0.54 | 0.58 | 0.62 |
| S Phase | Itr | 1.0 | 0.41 | 0.32 | 0.27 | 0.25 | 0.25 | 0.28 | 0.34 | 0.45 |
| | 10Itr | 1.0 | 0.34 | 0.27 | 0.20 | 0.16 | 0.16 | 0.20 | 0.29 | 0.37 |
| | Imax | 1.0 | 0.37 | 0.23 | 0.17 | 0.14 | 0.13 | 0.21 | 0.30 | 0.33 |
| | Itr | 0.5ind | 0.44 | 0.37 | 0.28 | 0.27 | 0.27 | 0.31 | 0.36 | 0.44 |
| | 10Itr | 0.5ind | 0.58 | 0.52 | 0.49 | 0.47 | 0.46 | 0.49 | 0.53 | 0.60 |
| | Imax | 0.5ind | 0.64 | 0.58 | 0.59 | 0.54 | 0.55 | 0.57 | 0.58 | 0.64 |
| T Phase | Itr | 1.0 | 0.42 | 0.30 | 0.22 | 0.19 | 0.20 | 0.22 | 0.30 | 0.42 |
| | 10Itr | 1.0 | 0.37 | 0.25 | 0.19 | 0.16 | 0.14 | 0.22 | 0.25 | 0.37 |
| | Imax | 1.0 | 0.33 | 0.29 | 0.20 | 0.13 | 0.15 | 0.22 | 0.30 | 0.34 |
| | Itr | 0.5ind | 0.45 | 0.36 | 0.29 | 0.28 | 0.28 | 0.28 | 0.36 | 0.45 |
| | 10Itr | 0.5ind | 0.64 | 0.57 | 0.54 | 0.52 | 0.51 | 0.52 | 0.57 | 0.60 |
| | Imax | 0.5ind | 0.67 | 0.66 | 0.61 | 0.59 | 0.59 | 0.62 | 0.63 | 0.68 |

First Issued: 2015-05-12 Latest Issue: 2017-05-08 Expiry Date: 2025-05-11

Page: 3 of 6


This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/632148

Schedule of Approval

Seal Locations & Photo showing Markings





First Issued: 2015-05-12 Latest Issue: 2017-05-08 Expiry Date: 2025-05-11

Page: 4 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/632148

Schedule of Approval

Product Variant Identification Details

| | METER TYPE CODE TABLE | | | | | |
|----------------------------|---------------------------|---------------------------|--|-------------------|--------------------------|-----------------------------|
| PRODUCT CODE | METERING | CURREN T | TYPE OF CONNECTION | METERING CLASS | COMMUNICATIO N OUTPUT | MODEL |
| VEM - | X | XXXXX | X | X | X | X |
| VİKO ENERGY METERING | M = MONO- PHASE | 1080 | D = DIRECT CONNECTION | A = CLASS A | 0= NONE | NULL CHARACTER NORMAL |
| | T = THREE- PHASE | 10100 | C = WITH CURRENT TRANSFORMER | B = CLASS B | 1 = RS232 | P= PREPAID |
| A = MODEL | C = COMBO (INDUSTRIAL) | x5 | V = WITH CURRENT/VOLTAGE TRANSFORMER | C = CLASS C | 2 = RS485 | |
| | | 1060 5 (80) 5 (100) | | | 3 = PLC | |

Marking

The following data shall be quoted on the meter name-plate:

- Identification of manufacturer (name or trade mark);
- Type designation;
- "CE" marking and supplementary metrology marking;
- EC-type examination certificate number;
- Serial number and year of production;
- Class index;
- Specified operating temperature range;
- Type of distribution network;
- Reference voltage;
- Reference (rated) current;
- Maximum current;
- Minimum current;
- Reference frequency;
- Constant of meter;
- Sign of double-square for meter of protective class II

First Issued: 2015-05-12 Latest Issue: 2017-05-08 Expiry Date: 2025-05-11

Page: 5 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.

No. 0086/BSI/632148

Schedule of Approval

Manufacturer Address: Panasonic Eco Solutions Elektrik Sanayi Ticaret Anonim Sirketi

Abdurrahmangazi Mah Ebubekir Cad. No: 44 34887

Samandıra Istanbul Turkey

Document Revision History

| Issue | Date | Comments |
|-------|------------|--|
| 1 | 12/05/2015 | Initial Issue. |
| 2 | 08/11/2016 | Minor hardware Changes, Software version updated. BSI project Ref 8624843. |
| 3 | 08/05/2017 | Company Name Change |

Conditions of Certification

- i) This 0086/BSI/632148 certificate remains valid unless cancelled or revoked, provided the conditions listed below are complied with and the equipment remains satisfactory in service.
- ii) This certificate must be used in conjunction with a certificate covering the product certification as required by Annex D or Annex F.
- iii) Detailed User instructions are to be provided with each product.
- iv) If the specified standards are amended during the validity of this certificate, this product type is to be re-approved.
- v) Any modifications to the Meter(s) listed on page 1 must be notified to BSI to confirm the Meter(s) continuing compliance to the relevant pattern approval standard(s).

First Issued: 2015-05-12 Latest Issue: 2017-05-08 Expiry Date: 2025-05-11

Page: 6 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.