

EU - Type Examination Certificate

This is to certify that:

Panasonic Eco Solutions Elektrik Sanayi
Ticaret Anonim Sirketi
Abdurrahmangazi mah
Ebubekir cad. No:44 Sancaktepe
Istanbul
34887
Turkey

Holds Certificate Number: 0086/BSI/668762

In respect of:

Annex : MI-003 – Active Electrical Energy Meter

Product : VEM-M580DB0T, 1 Phase 230V, 0.25-5(80)A, 50Hz, Class B, 3K7, IP54. Software version 1.02.

Specified Standard: EN 50470-1:2006, EN 50470-3:2006

on the basis that BSI carried out the relevant Type Examination procedures for the equipment identified above which was found to be in compliance with the Directive 2014/32/EU of the European Parliament and Council relating to the Measuring Instruments Directive (MID), subject to any conditions in the schedule attached hereto. The attached schedule of approval forms part of this certificate.

For and on behalf of BSI, a Notified
Body for the above Directive
(Notified Body Number 0086):



Gary Fenton, Global Product Certification Director

First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 1 of 6

...making excellence a habit.™

EU - Type Examination Certificate

No. 0086/BSI/668762

Schedule of Approval

Technical Data

Manufacturer	Panasonic	Meter Type	VEM-M580DB0T
Voltage Rating (Un)	230V	Current Rating (Imin-Iref(I _{max}))	250mA - 5A - 80A
Frequency (Fn)	50Hz	Active Accuracy Class (kWh)	Class B
Type of Circuit	1p2w	Temperature Range	-40 +70
Software Version No. Identification Location Checksum PCB/Schematic Ref	Ver 1.02 LCD: 0xB5DA 265136 R0.3 - 21/09/15	Bill of Material Nos.	92903701 dated 19/10/16
IP Rating	IP54	Insulation Protective Class	Class II
LED Pulse Constant	3.200 imp/kWh	Impulse Voltage Rating	6kV
AC Voltage Rating	4kV	Seal for EC Initial Verification Type	Plastic Seal on Top Cover and Seal on Terminal cover
Main Cover Seal Type	Push Fit Plastic seal	Integrity of Meter	Inaccessible without breaking the seal
Intended Location of the Meter	Outdoor	Type of Register	8 Digit LCD

Certification is based on the Turkey accredited laboratory ELDAS, Report number EL-301 R1 dated 13/10/2014, CMI Test Report number 6011-PT-TS024-14 dated 13/04/2015, NMI Report 1900849-01 dated 29/03/2017 and Technical Documentation supplied to BSI, project Ref 8689042.

Condition for Conformity Assessment according to Annex D or F

The influence factors for temperature, frequency and voltage, which are necessary to perform the conformity assessment according to annex D of F, are presented on page 3 of this certificate. Based on WELMEC Guide 11.1 Section 2.5.6, the sum of the square values is presented.

During the type approval examination the influence factors for temperature, frequency and voltage are determined per load point. The table below presents the sum of the square values per load, determined via the following formula

$$\delta e(T, U, f) = \sqrt{\delta e^2(T, I, \cos\varphi) + \delta e^2(U, I, \cos\varphi) + \delta e^2(f, I, \cos\varphi)}$$

Where:

$\delta e(T, I, \cos\varphi)$ = the additional error due to the variation of the temperature at a certain load.

$\delta e(U, I, \cos\varphi)$ = the additional percentage error due to the variation of the voltage at the same load.

$\delta e(f, I, \cos\varphi)$ = the additional percentage error due to the variation of the frequency at the same load.

First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 2 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).


EU - Type Examination Certificate

No. 0086/BSI/668762

Schedule of Approval

			Influence Factor for Temperature, Frequency & Voltage							
	Current	PF Cos	-40°C	-25°C	-10°C	+5°C	+30°C	+40°C	+55°C	+70°C
Single Phase	Imin	1.0	0.62	0.59	0.58	0.58	0.58	0.58	0.59	0.60
	Itr	1.0	0.28	0.18	0.12	0.07	0.06	0.09	0.14	0.20
	10Itr	1.0	0.30	0.26	0.22	0.22	0.21	0.23	0.24	0.25
	Imax	1.0	0.27	0.19	0.14	0.11	0.12	0.14	0.18	0.21
	Itr	0.5ind	0.45	0.40	0.39	0.38	0.38	0.39	0.41	0.43
	10Itr	0.5ind	0.27	0.24	0.19	0.18	0.18	0.19	0.20	0.26
	Imax	0.5ind	0.37	0.31	0.30	0.28	0.28	0.30	0.32	0.33
	Itr	0.8cap	0.33	0.28	0.26	0.24	0.24	0.25	0.25	0.27
	10Itr	0.8cap	0.31	0.24	0.21	0.17	0.16	0.17	0.19	0.23
	Imax	0.8cap	0.37	0.29	0.23	0.21	0.20	0.21	0.24	0.27

Seal Locations



First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 3 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU - Type Examination Certificate

No. 0086/BSI/668762

Schedule of Approval

Photo



Marking

The following data shall be quoted on the meter name-plate:

- Identification of manufacturer (name or trade mark & address);
- Type designation;
- "CE" marking and supplementary metrology marking followed with year of manufacture;
- EU-type examination certificate number;
- Serial number;
- Class index;
- Specified operating temperature range;
- Type of distribution network;
- Reference voltage;
- Reference (rated) current;
- Maximum current;
- Minimum current;
- Reference frequency;
- Constant of meter;
- Sign of double-square for meter of protective class II.

First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 4 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU - Type Examination Certificate

No. 0086/BSI/668762

Schedule of Approval

Product Variant Identification Details

METER TYPE CODE TABLE						
PRODUCT CODE	METERING	CURRENT	TYPE OF CONNECTION	METERING CLASS	COMMUNICATION OUTPUT	MODEL
VEM -	X	XXXXX	X	X	X	X
VİKO ENERGY METERING A = MODEL	M = MONO-PHASE	1080	D = DIRECT CONNECTION	A = CLASS A	0= NONE	NULL CHARACTER NORMAL
	T = THREE-PHASE	10100	C = WITH CURRENT TRANSFORMER	B = CLASS B	1 = RS232	P= PREPAID
	C = COMBO (INDUSTRIAL)	x5	V = WITH CURRENT/VOLTAGE TRANSFORMER	C = CLASS C	2 = RS485	
		1060 5 (80) 5 (100)			3 = PLC	

Manufacturer Address:

Panasonic Eco Solutions Elektrik Sanayi Ticaret Anonim Sirketi
Abdurrahmangazi Mah
Ebubekir Cad. No: 44 34887
Samandıra
Istanbul
Turkey

Document Revision History

Issue	Date	Comments
1	08/05/2017	Initial Issue.

First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 5 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU - Type Examination Certificate

No. 0086/BSI/668762

Schedule of Approval

Conditions of Certification

- i) This 0086/BSI/668762 certificate remains valid unless cancelled or revoked, provided the conditions listed below are complied with and the equipment remains satisfactory in service.
- ii) This certificate must be used in conjunction with a certificate covering the product certification as required by Module D or Module F.
- iii) Detailed User instructions are to be provided with each product.
- iv) If the specified standards are amended during the validity of this certificate, this product type is to be re-approved.
- v) Any modifications to the Meter(s) listed on page 1 must be notified to BSI to confirm the Meter(s) continuing compliance to the relevant pattern approval standard(s).



First Issued: 2017-05-08

Latest Issue: 2017-05-08

Expiry Date: 2027-02-07

Page: 6 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.