



## Food and Environmental Proficiency Testing Unit

The final date for return of results is indicated on the accompanying request/report form

### Contact details:

The Organisers - FEPTU  
Public Health England  
61 Colindale Avenue,  
London, NW9 5EQ, UK.  
Fax: +44 (0) 20 8200 8264  
Tel: +44 (0) 20 8327 7119  
e-mail: foodeqa@phe.gov.uk

Results to be returned via an online survey as  
per email sent to your laboratory

<https://www.gov.uk/government/collections/external-quality-assessment-ega-and-proficiency-testing-pt-for-food-water-and-environmental-microbiology#mycobacterium-spp.-in-water-scheme>

## ***Mycobacterium* spp. in Water Scheme – Instruction Sheet**

*The LENTICULE® discs are provided in screw-cap plastic vials (with a desiccant). Each LENTICULE disc rests on the filter support containing the desiccant. The **desiccant** should be orange in colour. Please contact us if this is not the case. Each LENTICULE disc requires reconstitution by a process of rehydration and dispersion prior to examination, as described below.*

Refer also to the Safety Data Sheet: [www.gov.uk/government/publications/safety-data-sheet-for-lenticules](http://www.gov.uk/government/publications/safety-data-sheet-for-lenticules)

### Storage:

- Store the samples at **-20 ± 5°C** on receipt.
- Allow the LENTICULE discs to reach ambient temperature (5 - 10 minutes) **before** reconstituting in diluent.

### Reconstitution:

- Open the sample container and transfer the LENTICULE disc into approximately **9mL** 0.1% peptone saline (maximum recovery diluent (MRD)) that has been allowed to reach ambient temperature.
- Leave to stand at ambient temperature for 10 -12 minutes to rehydrate.
- Tighten the cap of the bottle and shake to disperse the micro-organisms.
- Transfer all the inoculated MRD to a sample bottle containing 1L sterile deionised or distilled water at ambient temperature.
- Rinse with approximately 2mL from the sample bottle, ensuring all liquid is transferred back to the 1L sample.
- Disperse the inoculum by inverting approximately 30 times.

### Examination:

- Each reconstituted sample is equivalent to 1L water.
- Undertake the sample examinations between 30 - 45 minutes of reconstitution.
- Examine in accordance with your routine procedures.

### Return of Results:

- Return your results on-line using the web-based reporting system.
- The distribution will be closed on the deadline date for return of results.
- Results cannot be accepted after the distribution has closed.