



**Virus proficiency testing scheme –
Preliminary Intended Results Notification**

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The quantification results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **NHV007** Samples: **NHV0013 and NHV0014**

Intended results:

		NHV0013	NHV0014
Virus	GI	Positive	Negative
	GII	Positive	Negative
	HAV	Negative	Positive
Quantity (copies/per sample)	GI	3.4x10 ⁴	-
	GII	5.5x10 ²	-
	HAV	-	1.4x10 ³

qRT-PCR was carried out using the RNA UltraSense™ One-Step Quantitative RT-PCR System (ThermoFisher), on a Stratagene Mx3005P real-time PCR machine

Laboratory identification numbers of participants who did not return results this distribution:
744, 1267, 2135 and 2634

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched shortly. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. **However please note due to suspension of our services from April 2020 onwards, these will not be dispatched until services have been resumed.**

We will confirm the next distribution of the Virus testing scheme.

Validated by:

Zak Prior
Project lead & general
laboratory manager

Date: 23 March 2020