

Protecting and improving the nation's health

Drinking Water Scheme – Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: W193 Samples: W193A and W193B

Intended results:

Enumeration	W193A Median	W193B Median
Coliform bacteria per 100mL	36	76
Escherichia coli per 100mL	23	43
Enterococci per 100mL	49	0
Pseudomonas aeruginosa per 100mL	15	31
Clostridium perfringens per 100mL	40	0
Colony count 37°C/48 h per mL	40	54
Colony count 22°C/72 h per mL	1.1x10 ²	52

Sample contents:

W193A	Enterobacter cloacae, E. coli, Enterococcus faecalis, P. aeruginosa, C. perfringens, Pseudomonas putida and Micrococcus sp.	
W193B	Klebsiella pneumoniae, E. coli, P. aeruginosa, Clostridium sordellii and Staphylococcus aureus	

Laboratory identification numbers of participants who did not return results this distribution:

18, 89, 464, 1094, 1120, 1258, 1267, 1563, 1881, 1956, 2344 and 2652

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Drinking Water Scheme samples will be dispatched during the week beginning 28 September 2020.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:

Lili Tsegaye

Lili Tsegaye
Laboratory Supervisor

Date: 01 September 2020



0006