

Protecting and improving the nation's health

## **Recreational and Surface Water Scheme** River, Lake and Stream Water Samples Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: S96 Samples: S96A and S96B

## Intended results:

Enumeration	S96A Median	S96B Median
Coliform bacteria in 100mL	60	41
Escherichia coli in 100mL	57	28
Enterococci in 100mL	66	96
Faecal coliforms in 100mL	57	28
Clostridium perfringens in 100mL	0	38
Salmonella spp.	Not detected	Detected

## Sample contents:

S96A	Enterobacter cloacae, E. coli, Enterococcus faecalis, Pseudomonas aeruginosa and Staphylococcus haemolyticus	
S96B	Klebsiella oxytoca, E. coli, Enterococcus durans, Clostridium perfringens and Salmonella Enteritidis	

Laboratory identification numbers of participants who did not return results this distribution: 854, 1088, 1267, 1881 and 2592

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Recreational and Surface Water Scheme samples will be Swimming pool waters and will be dispatched during the week beginning 28 September 2020.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:

Margaret Nienga Laboratory Supervisor Date: 01 September 2020