



Legionella Isolation Scheme Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **G119** Samples: **G119A and G119B**

Intended results:

	G119A	G119B
*cfu L ⁻¹	5.4x10 ⁴	6.4x10 ²
Species	<i>Legionella jordanis</i>	<i>Legionella pneumophila</i>
Serogroup	-	Serogroup 1

* colony forming units per litre

Sample contents:

G119A	<i>Legionella jordanis</i> and <i>Staphylococcus epidermidis</i>
G119B	<i>Legionella pneumophila</i> , <i>Escherichia coli</i> and <i>Staphylococcus saprophyticus</i>

Laboratory identification numbers of participants who did not return results this distribution:
979, 1009, 1019, 1159, 1213, 1267, 1563, 1773, 1956, 1971, 2077, 2114, 2328, 2605 and 2651

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Legionella Isolation Scheme samples will be dispatched during the week beginning 28 September 2020.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:


Margaret Njenga
Laboratory Supervisor

Date: 7 September 2020



0006