



Drinking Water Scheme – Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **W194** Samples: **W194A and W194B**

Intended results:

Enumeration	W194A Median	W194B Median
Coliform bacteria per 100mL	46	26
<i>Escherichia coli</i> per 100mL	0	26
Enterococci per 100mL	34	57
<i>Pseudomonas aeruginosa</i> per 100mL	0	27
<i>Clostridium perfringens</i> per 100mL	58	35
Colony count 37°C/48 h per mL	17	1
Colony count 22°C/72 h per mL	17	2

Sample contents:

W194A	<i>Klebsiella pneumoniae</i> , <i>Enterococcus faecalis</i> , <i>Clostridium perfringens</i> and <i>Staphylococcus sciuri</i>
W194B	<i>E. coli</i> , <i>Enterococcus faecalis</i> , <i>Pseudomonas aeruginosa</i> and <i>Clostridium perfringens</i>

Laboratory identification numbers of participants who did not return results this distribution:

1366, 1654, 2294, 2416, 2544, 2651 and 2677


The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Drinking Water Scheme samples will be dispatched during the week beginning 15 February 2021.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:


Margaret Njenga
Laboratory Supervisor

Date: 2 November 2020



0006