

## Guidance for manufacturers on reporting adverse incidents involving Inferior Vena Cava (IVC) Filters under the vigilance system

To be read in conjunction with the guidelines on a medical devices vigilance system [MEDDEV 2.12/1](#)

Report as individual events (in line with MEDDEV timescales)	Can be included in periodic summary reports (PSR)*		Report at the time the adverse trend is identified	Generally not reportable
<ul style="list-style-type: none"> <li>• Death that is probably or possibly device-related</li> <li>• Complications during deployment or placement eg:               <ul style="list-style-type: none"> <li>- premature release</li> <li>- partial or incomplete expansion</li> <li>- deformation (such as crossed or twisted legs or arms)</li> <li>- asymmetric deployment (malapposition)</li> </ul> </li> <li>• Partial or multiple fractures</li> <li>• Device migration/secondary movement with or without embolization</li> <li>• Recurrent or fatal pulmonary embolism</li> </ul>		<b>Periodicity</b>	<ul style="list-style-type: none"> <li>• IVC wall penetration &lt;3mm</li> <li>• Vascular access and device placement related problems eg:               <ul style="list-style-type: none"> <li>- device misplacement / improper placement</li> <li>- pneumothorax</li> <li>- air embolism</li> <li>- haematoma / bleeding / haemorrhage</li> <li>- intimal tear</li> </ul> </li> <li>• Inferior vena cava thrombosis / occlusion / restriction of blood flow through the filter or venous insufficiency</li> <li>• Systemic Infection</li> <li>• Adverse reaction</li> </ul>	<ul style="list-style-type: none"> <li>• Death if there is evidence that it is <b>not</b> device related</li> <li>• Access site thrombosis or stenosis</li> <li>• Infection at puncture site</li> </ul>
	IVC wall perforation / erosion / penetration >3mm eg:	3 monthly		
	Retrieval difficulties / failure to retrieve	3 monthly		
	Progressive tilting / angulation	6 monthly		

\* If you can't use PSR, then report these events individually.