

## Guidance for manufacturers on reporting adverse incidents involving Neurostimulators under the vigilance system

To be read in conjunction with the guidelines on a medical devices vigilance system [MEDDEV 2.12/1](#)

Report as individual events (in line with MEDDEV timescales)	Report as periodic summary reports (PSR)*		Report at time adverse trend is identified
Telemetry failure (during or after implantation)		<b>Periodicity</b>	A change in stimulation due to fibrosis around electrode
Fracture of accessory / introducer items	Lead fracture	3 monthly	
Premature battery depletion	Lead shorts	3 monthly	
Inability of rechargeable battery to hold charge	Lead migration	3 monthly	
Mechanical failure of pulse generator (including internal components)	High lead impedance	3 monthly	
Electrical failure of pulse generator	Electromagnetic interference	3 monthly	
Unexplained loss of therapy			
Neuropraxia			
Tissue Damage			

\* If you can't use PSR, then report these events individually.