



## Endoscope Rinse Water Scheme Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **EW26**      Samples: **EW26A and EW26B**

### Intended results:

	EW26A		EW26B	
Enumeration	Median	Conclusion*	Median	Conclusion*
TVC per 100mL	45	Risk Assessment Required	64	Risk Assessment Required

\* Conclusions determined from: *Health Technical Memorandum 01-06: Decontamination of flexible endoscopes Part E: Testing methods*

### Sample contents:

<b>EW26A</b>	<i>Aspergillus fumigatus</i> and <i>Pseudomonas aeruginosa</i>
<b>EW26B</b>	<i>Pseudomonas aeruginosa</i> and <i>Saccharomyces cerevisiae</i>

Laboratory identification numbers of participants who did not return results this distribution:

157, 1647, 1735, 1881, 1956, 1971 and 2076

*The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.*

**Repeat samples** are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

*The next distribution of Endoscope Rinse Water Scheme samples will be dispatched during the week beginning 5 April 2021.*

*If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.*

Validated by:

  
Margaret Njenga  
Laboratory Supervisor

Date: 1 February 2021



0006