



Drinking Water Scheme – Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **W195** Samples: **W195A and W195B**

Intended results:

Enumeration	W195A Median	W195B Median
Coliform bacteria per 100mL	42	49
<i>Escherichia coli</i> per 100mL	34	27
Enterococci per 100mL	34	49
<i>Pseudomonas aeruginosa</i> per 100mL	88	0
<i>Clostridium perfringens</i> per 100mL	0	49
Colony count 37°C/48 h per mL	3	79
Colony count 22°C/72 h per mL	3	82

Sample contents:

W195A	<i>Klebsiella oxytoca</i> , <i>E. coli</i> , <i>Enterococcus durans</i> , <i>P. aeruginosa</i> and <i>Chryseobacterium indologenes</i>
W195B	<i>Enterobacter cloacae</i> , <i>E. coli</i> , <i>Enterococcus hirae</i> , <i>Clostridium perfringens</i> and <i>Staphylococcus xylosus</i>

Laboratory identification numbers of participants who did not return results this distribution:
1097, 1114, 1425, 1507, 1585, 2416, 2478 and 2651


The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Drinking Water Scheme samples will be dispatched during the week beginning 5 April 2021.

If you require any technical advice or wish to make comments or suggestions relating to the scheme, please contact us.

Validated by:


Margaret Njenga
Laboratory supervisor

Date: 22 March 2021



0006