

EC Type Examination Certificate Number: **UK/ 0120/ SGS0066**

Itron Metering Solutions UK Ltd

PO Box No.3,
Talbot Road,
Stretford,
Manchester
M32 0XX

Instrument Identification:
U65

G40 diaphragm gas meter

Instrument Traceable Number
UK/ 0120/ SGS0066

EC Directive 2004/22/EC **on Measuring Instruments Annex B**

It is certified that the manufacturer's technical design and specimen for the above instrument has been examined and, based on the evidence submitted, it is considered that the instrument conforms to the requirements of MI-002 of EC Directive 2004/22/EC.

This certificate is to be used in conjunction with a certificate covering the product verification as required in Annex D or Annex F

This certificate is valid until 26th September 2020
Issue 2

Certification is based on report number(s) MID-B- MID-B-187672-U65Ra dated 9th June 2016

Notified Body Number 0120

Authorised Signature



CE
0120

Notified Body
SGS United Kingdom Ltd, 202B Worle Parkway, Weston-super-Mare, BS22 6WA UK
t +44 (0)1934 522917 f +44 (0)1934 522137 www.sgs.com
Contact Address
SGS United Kingdom Ltd, Unit 10, South Industrial Estate, Bowburn, Durham, DH6 5AD UK
t +44 (0)191 377 2000 f +44 (0)191 377 2020 www.sgs.com

EC Type Examination Cert.

Approval Annex for EC Type Examination Cert. No. UK/ 0120/ SGS0066


Manufacturer	:	Itron Metering Solutions UK Ltd.	
Type of Meter(s)	:	U65	
Description of Meter	:	G40 diaphragm gas meter.	
Accuracy Class (es)	:	1.5	Software Version No : Not Applicable.
Cyclic volume:		25dm ³	U65 Bill Of Materials No : DD-032-0B
Meter Location	:	Indoor, or in suitable enclosure.	Sealing Type : Embossed plastic seal. Location : Pulse Ready- Fitted to top of index cover. Cyble-- Fitted to bottom of index Cover.
Integrity of Meter	:	Inaccessible without breaking seal.	Temperature Range : -10°C to +40°C
Range of Gases	:	Second Family Group-IIA	Maximum Operating Pressure : 75mbar
Flow Range	:	Qmax 65m ³ /h Qmin 0.4m ³ /h	Orientation : Connection ports vertical

The meter is approved with either the Pulse Ready or Cyble Index

Photograph Showing Complete U65 25dm³ Meter with Pulse Ready Index




Pulse Ready Meter Index



Pulse Ready Meter Index showing blue security seal location.
The index is shown without the optional 45-0000 pulse module fitted

Detail of Seal Location on the Pulse Ready Index



Pulse Ready Seal Detail



The Itron seal symbol and date will be heat embossed on the Index seal X300078 after the meter has been certified.

Optional Pulse Output Module 45-0000 for the Pulse Ready Index



The Module provides a 100dm³ pulse VFC output. The module is fitted to the Index with a screw which may be covered with a seal.


Cyble Meter Index



Detail of the Seal and its Location on the Pulse Cyble Index



Top View of Seal with the impressed Itron 'O' Symbol.



Side View of Seal



Location Hole for Metrological Seal in the Cyble Index



UK/ 0120/ SGS0066 9 th June 2016 Issue 2
--

Approved Product Variations:

Pulse output module 45-0000 may be attached to the meter index

Modifications to the meters described according to approval No. **UK/ 0120/ SGS0066** must be notified to the issuing body to confirm the meter's continuing compliance to the relevant pattern approval standard(s). Modifications to the pulse output module 45-0000 relating to the readability of the meter index and the applied load on the index, must also be agreed with the issuing body before being implemented.