

Type examination certificate UK/0126/0314

Issued by:

NMO Approved Body Number 0126

In accordance with the requirements of the Measuring Instruments Regulations 2016 (S.I. 2016 No. 1153), as amended, this type examination certificate has been issued to:

Flonidan A/S Islandsvej 29 8700 Horsens Denmark

In respect of a diaphragm gas meter designated the **Uniflo G4S** and having the following characteristics:

Accuracy class: 1.5 Mechanical measuring unit (either two pipe connection or coaxial connection): UG G4 Maximum flow rate: 6 m³/h Cyclic volume: 1.2 dm³

The necessary data (principal characteristics, alterations, securing, functioning etc) for identification purposes and conditions (when applicable) are set out in the Descriptive Annex to certificate DK-0200-MI002-013 Revision 12 (issued on 26 May 2020).

The conditions of use are set out in the Annex to this certificate.

Issue date: 03 November 2021 Valid until: 04 November 2029

Grégory Glas Lead Technical Manager For and on behalf of the Head of NMO



Annex – Conditions of use

1 References

All EU references in the Descriptive Annex to certificate DK-0200-MI002-013 Revision 12 shall be read in this certificate as their equivalent in the Measuring Instruments Regulations 2016 (S.I. 2016 No. 1153), as amended.

2 Parts / Evaluation / Test certificates

References to Parts / Evaluation / Test certificates in the Descriptive Annex to certificate DK-0200-MI002-013 Revision 12 shall be replaced with references to the equivalent NMO certificates.

3 Technical documentation

The instrument is fully described in the technical file held at NMO.

4 Markings and inscriptions

The markings and inscriptions shall fulfil the requirements of Part 4, Chapter 4 of the Measuring Instruments Regulations 2016 (S.I. 2016 No. 1153), as amended.

5 Sealings

The sealings must bear an official mark, which may be either:

- a mark of the manufacturer and/or manufacturer's representative, or
- an official mark of a verification officer.

Certificate history

Issue No.	Date	Description
UK/0126/0314	03 November 2021	Type examination certificate first issued.
-	-	No revisions have been issued.