





## YOFULY Vacuum Therapy Machine FX024D

Aspect	Details
Images	 
Alert Number	2212-0001
Product Type	Electrical appliances and equipment - Vacuum Therapy Machine
Product Identifiers	Model number: BL-600 ASIN: B07Y8B1S6Q (Please note the upgraded model currently for sale with this ASIN number is not affected by this safety report). SKU: X0014KYRJD Other product identifiers: FBA15G6PSSZ7U000011
Product Description	Vacuum therapy suction machine
Country of Origin	China
Counterfeit	Unknown
Risk Type	Fire and burns
Risk Level	High
Risk Description	The product presents a high risk of fire and burns as when tested the plug does not comply with the relevant standard (BS1363) as it fitted with an unsafe fuse which has no granular filler in the fuse itself. If a fault were to occur, the plug could explode from the wall causing a fire, causing burns to the consumer from touching the product or debris. Additionally, the fuse does not comply with the dimensional requirements of the relevant standard. The internal wiring is insulated in basic insulation, and they were in contact with accessible conductive parts. The product does not comply with the marking requirements of the relevant standard and the instructions are considered inadequate. The product does not meet the requirements of the Electrical Equipment (Safety) Regulations 2016.
Corrective Measures	The import has been rejected at the border.
Online Marketplace	The affected product listing has been removed by the online marketplace (Amazon).
Notifier	Local Authority Trading Standards