




<b>BEHA-AMPROBE MULTIMETERS</b>	
<b>Aspect</b>	<b>Details</b>
<b>Images</b>	
<b>Alert Number</b>	2303-0015
<b>Product Type</b>	Electrical appliances and equipment - Multimeter
<b>Product Identifiers</b>	Brand: BEHA-AMPROBE Affected model numbers: AM-500-EUR, AM-510-EUR, HEX60-D, AM-500-EUR KIT, AM-510-EUR KIT. Affected serial numbers: Within the serial number range of 535900001 – 587799999 and any other serial number that starts with any number other than a 5.
<b>Product Description</b>	The products are red/black/grey digital multimeters for making electrical measurements.
<b>Country of Origin</b>	China
<b>Counterfeit</b>	Genuine
<b>Risk Type</b>	Fire
<b>Risk Description</b>	<p>The product has been recalled as it was discovered that an improperly rated fuse for CAT III 600V applications was installed in the affected products. The breaking capacity of the installed fuse is insufficient for the prospective fault current under foreseeable misuse conditions (abnormal application of electrical installation voltage above 250 volts between A and COM terminals) and may lead to venting or rupture of the fuse and the subsequent possibility of arc explosion. An arc explosion within the product enclosure may readily escape and result in a fire or severe bodily harm.</p> <p>The product does not comply with The Electrical Equipment (Safety) Regulations 2016</p>
<b>Corrective Measures</b>	Recall of the product from end users. For more information, please see the BEHA-AMPROBE recall site here: <a href="https://www.beha-amprobe.com/en/announcements/safety-warning-please-readimportant-product-safety-recall-information-regarding">https://www.beha-amprobe.com/en/announcements/safety-warning-please-readimportant-product-safety-recall-information-regarding</a>
<b>Online Marketplace</b>	N/A
<b>Notifier</b>	Local Authority Trading Standards