



Fluke - 8x V Series- Digital Multimeters (DMM)	
Aspect	Details
Images	
Alert Number	2301-0135
Product Type	Electrical appliances and equipment - Digital Multimeters (DMM)
Product Identifiers	<p>Within the Fluke 8x V Series, the following models:</p> <ul style="list-style-type: none"> Fluke 87V Fluke 87vMax Fluke 83V Fluke 88V <p>The affected serial numbers are as follows:</p> <p>All Fluke 83V, 87V, 87V MAX or Fluke 88 V models with a serial number between 46280001 and 55370001 (both included).</p> <p>Users can check if they have one of the affected appliances on the back of the appliances. Here you will find the serial number</p> 
Product Description	The product is used to make electrical measurements and tests including AC and DC voltage, temperature (thermocouple), resistance, continuity, capacitance, AC and DC current by professional operators.
Country of Origin	United States
Counterfeit	No
Risk Type	Electric Shock
Risk Level	High

Risk Description	<p>Some products may have an intermittent electrical connection in the COM terminal which may cause an operator to incorrectly assess a hazardous electrical installation as being disconnected/safe to touch.</p> <p>This incorrect assessment could result in electric shock/electrocution/burns due to arc flash if contact is made with the installation conductors.</p>
Corrective measures	<p>The importer has withdrawn the affected product from the market and informed all distributors to stop sale and return any affected products. The importer has informed and contacted distributors/end customers with information related to the potential safety risk and provided instructions to perform a simple test and return the DMMs for repair, free of charge.</p> <p>The importer has issued a safety notice on its website to all end-customers who have purchased the products.</p> <p>Important Safety Notice: Fluke 8x V Series Digital Multimeters Fluke</p>
Online Marketplace	N/A
Notifier	Health and Safety Executive

The OPSS Product Safety Alerts, Reports and Recalls Site can be accessed at the following link: <https://www.gov.uk/guidance/product-recalls-and-alerts>