



Medicines & Healthcare products
Regulatory Agency

10 South Colonnade
Canary Wharf
London
E14 4PU
United Kingdom
[gov.uk/mhra](https://www.gov.uk/mhra)

[REDACTED]

29 August 2023

FOI 23/607

Dear [REDACTED]

Thank you for your additional Freedom of Information request dated 15th August 2023, where you asked for "information on adverse incidents relating to devices that deliver drugs to the brain". Including the following GMDN codes: 61854, 61376, 10691 and 61798.

Further to your request, please see Table 1 which includes the number of UK, spontaneous adverse incident reports associated with the above mentioned GMDN codes and the reported device problem, in a similar format to the data previously provided. Additionally, please see Table 2 which displays the number of UK, spontaneous adverse incident reports associated with the GMDN codes and the reported clinical effect. It is important to note that a singular report can contain more than one device problem and associated clinical effect.

Please be aware that where less than 5 reports have been received the number has not been included within the table. This is in compliance with data protection laws and to protect patient/reporter confidentiality.

Please note that the inclusion of a report on our adverse incident database does not necessarily mean the events described were caused by that device but could be due to unrelated patient/user factors. In addition, details of the reports may have changed since the report was submitted.

Table 1. The number of adverse incidents reported in association with the respective drug delivery system GMDN code, including reported medical device problem.

Device Type	GMDN Code	Reported Medical Device Problem(s)	Number of Reports
Intracerebral cannula, non-implantable, reusable	61854	-	0
Intracerebral infusion/drainage catheter, long-term	61376	Inaccurate Information	<5
Vascular microcatheter	10691	Adverse Event Without Identified Device or Use Problem	<5
		Mechanical Problem	<5
		Physical Resistance/Sticking	<5
		Material Split, Cut or Torn	<5
		Fracture	<5
		Detachment of Device or Device Component	6
		Break	<5
		Material Fragmentation	<5
		Component Missing	<5
		Material Separation	<5
Ventricular cerebrospinal fluid drainage catheterization kit	61798	Insufficient Information	<5
		Fracture	<5
		None reported	25

Table 2. The number of adverse incidents reported in association with the respective drug delivery system GMDN code, including reported clinical effect.

Device Type	GMDN Code	Reported Clinical Effect(s)	Number of Reports
Intracerebral cannula, non-implantable, reusable	61854	-	0
Intracerebral infusion/drainage catheter, long-term	61376	None reported	<5
Vascular microcatheter	10691	Device Embedded in Tissue or Plaque	<5
		Low Blood Pressure/Hypotension	<5
		Foreign Body in Patient	<5
		Pseudoaneurysm	<5
		Perforation of Vessels	<5
		Stroke/Cardiovascular accident (CVA)	<5
		Thromboembolism	<5
		Vascular Dissection	<5
		Vasoconstriction	<5
		Insufficient Information	<5
		No Clinical Signs, Symptoms or Conditions	9
		None reported	<5
Ventricular cerebrospinal fluid drainage catheterization kit	61798	No Clinical Signs, Symptoms or Conditions	<5
		None reported	25

The data must be read together with the following explanations:

- The majority of reports indicate an issue experienced by a single user. However, some cases may represent the same user experiencing further issues.
- Reports do not necessarily represent an individual patient. Individuals may report an incident at any time after the event and people can make multiple reports at any time after the use of the device and on the same issue. Where possible, multiple reports for

the same event are linked. However, as reporters are not required to complete all fields, we cannot always be sure enough to link every duplicate.

- It should be noted that this information may include a range of recognised complications related to this type of procedure and does not necessarily indicate a fault with any particular device.
- When interpreting the above data it is important to note that the number of reports received should not be used as a basis for determining the incidence of a reaction as neither the total number of reactions occurring, nor the number of patients using the delivery device is known.
- The numbers may include reports where the incident has been taken from published literature.
- These numbers of reports are accurate at the time they are extracted from our database and minor changes in the numbers can occur if the reporter of the incident gives us more details later.
- Adverse incident reports by members of the public are voluntary but play a substantial part in the successful operation of the vigilance system. All reports received via Yellow Card are sent to the relevant manufacturer (if known and anonymised as appropriate) to feed into the vigilance system.
- Adverse incident reports include mandatory reporting by manufacturers to MHRA for certain types of incidents that occurred in the UK as part of the regulatory post market surveillance 'vigilance' system. The principal purpose of this system is to improve the protection of health and safety of patients. This is to be achieved by the evaluation of reported AIRs and, where appropriate, dissemination of information, which could be used to prevent such repetitions, or to alleviate the consequences of such adverse events.

The Freedom of Information Act only entitles you access to information – the information supplied is subject to Crown copyright, and there are some restrictions on its re-use. For information on the reproduction or re-use of MHRA information, please visit <https://www.gov.uk/government/publications/reproduce-or-re-use-mhra-information/reproduce-or-re-use-mhra-information>.

If you have a query about the information provided, please reply to this email.

If you are dissatisfied with the handling of your request, you have the right to ask for an internal review. Internal review requests should be submitted within two months of the date you receive this response and addressed to: info@mhra.gov.uk

Please remember to quote the reference number above in any future communications.

If you were to remain dissatisfied with the outcome of the internal review, you would have the right to apply directly to the Information Commissioner for a decision. Please bear in mind that the Information Commissioner will not normally review our handling of your request unless you have first contacted us to conduct an internal review. The Information Commissioner can be contacted at:

Information Commissioner's Office
Wycliffe House
Water Lane
Wilmslow
Cheshire

SK9 5AF

Yours sincerely,

FOI Team,
Safety and Surveillance group